



Certificate

Certificate no.

US 72091573 01

License Holder:

Aurora Mfg.
325 McGill Ave., Suite 195
Concord NC 28027
USA

Manufacturing Plant:

Aurora Mfg.
325 McGill Ave., Suite 195
Concord NC 28027
USA

Test report no.: USA-WS 30893462 002

Client Reference: Howie Morrison

Tested to: UL 60601-1:2003 R6.03

Certified Product: Fluidized Therapy Bed

License Fee - Units

Model Designation: HydroAire Model 3000

7

Rated Voltage: AC 120V, 60Hz

Rated Current: 12.0A

Protection Class: I

Patient Load: 300 lbs max.

Special Remarks: The applicable requirements of IEC 60601-2-38:1996+A1 were also taken into consideration.

7

Appendix: 1, 1-4

Inh. = 790997 / Deb. = 790997 / Fert. = 790997

Licensed Test mark:



Signature

Dipl.-Ing. M. Glagla
QA Certification Officer

Date of Issue
(day/mo/yr)

07/07/2009